

## Instrument Security Procedures

**Model:**

Fluke 73, 75, 77, 78

**Product Name:**

Digital Multimeter

**Instrument Description:**

Digital Multimeter

**Memory Description:**

There is no memory chip in these meters. The operating code in the unit is contained within Masked ROM. The volatile memory is contained in RAM and is cleared/lost when the function switch is changed and when the unit is turned off. This memory is not user accessible. The Masked ROM is a propriety Fluke part.

**Memory Cleaning Instructions:**

NA